

Ст. Добронос Дмитро Сергійович гр. ПІ-31

Анотація

Переддипломна практика проходила на підприємстві ДП «Укрметртестстандарт» у науково-виробничому відділі вимірювань сили, крутильного моменту сили та об'єму рідини. Підрозділ займається проведенням оцінки відповідності, повірки та калібрування засобів виміральної техніки за такими видами вимірювань: вимірювання механічних величин; вимірювання параметрів потоку, витрати, рівня та об'єму речовин. Практика почалася екскурсією, де були показані: установка еталонна для повірки моментних ключів УПМК 2000, еталонна силоримірювальна машина ОСМ- 2-5. Протягом практики був досліджений державний первинний еталон одиниці сили ДЕТУ 02-07-11. Згідно з індивідуальним завданням була проведена повірка динамометра.

Annotation

Pre-diploma practice was held at the enterprise SE "Ukrmetrtteststandart" in research and production department measuring force, torque force and fluid volume. The department is engaged in the conformity assessment, verification and calibration of measuring instruments for the following species measurement: measurement of mechanical quantities; flow measurement, flow, level and volume substances. The practice began excursion, where they were shown: appliance standards for verification torque wrenches UPMK 2000, reference force measuring machine OSM-2-5. During practice has studied the state primary standard power unit DETY 02-07-11. According to the individual tasks was performed calibration of the dynamometer.